

HACH METHOD 8195
DETERMINATION OF TURBIDITY BY NEPHELOMETRY

Revision 2.0
February 1999

Hach Company
5600 Lindbergh Drive
Loveland, Colorado 80538-8998

HACH METHOD 8195

DETERMINATION OF TURBIDITY BY NEPHELOMETRY

1.0 SCOPE AND APPLICATION

- 1.1 This method covers the determination of turbidity in drinking, ground, surface, and saline waters, domestic and industrial wastes.
- 1.2 The applicable range is 0 to 40 nephelometric turbidity units (NTU). Higher values may be obtained with dilution of the sample.

2.0 SUMMARY OF METHOD

- 2.1 The method is based upon a comparison of the intensity of light scattered by the sample under defined conditions with the intensity of light scattered by a standard reference suspension. The higher the intensity of scattered light, the higher the turbidity. Readings, in NTU'S, are made in a nephelometer designed according to specifications given in sections 6.1 and 6.2. Primary standard suspensions are used to calibrate the instrument. These primary standard suspensions are listed below:

2.1.1 Primary standards:

2.1.1.1 Formazin polymer is used as a primary turbidity suspension for water because it is more reproducible than other types of standards previously used for turbidity analysis.

2.1.2.2 StablCal™ formazin turbidity standards are available from Hach Company.

A secondary standard suspension is used as a daily calibration check and is monitored periodically for deterioration using one of the primary standards.

3.0 DEFINITIONS

- 3.1 CALIBRATION BLANK (CB) -- A volume of reagent water fortified with the same matrix as the calibration standards, but without the analytes, internal standards, or surrogates analytes.
- 3.2 INSTRUMENT PERFORMANCE CHECK SOLUTION (IPC) -- A solution of one or more method analytes, surrogates, internal standards, or other test substances used to evaluate the performance of the instrument system with respect to a defined set of criteria.
- 3.3 LABORATORY REAGENT BLANK (LRB) -- An aliquot of reagent water or other blank matrices that are treated exactly as a sample including exposure to all glassware, equipment, solvents, reagents, internal standards, and surrogates that are used with other samples. The LRB is used to determine if method analytes or other interferences are present in the laboratory environment, the reagents, or the apparatus.
- 3.4 LINEAR CALIBRATION RANGE (LCR) -- The concentration range over which the instrument response is linear.

- 3.5 MATERIAL SAFETY DATA SHEET (MSDS) -- Written information provided by vendors concerning a chemical's toxicity, health hazards, physical properties, fire, and reactivity data including storage, spill, and handling precautions.
- 3.6 PRIMARY CALIBRATION STANDARD (PCAL) -- A suspension prepared from the primary dilution stock standard suspension or commercially available standards (StablCal™ formazin turbidity standard and AEPA-1) that have shown to be equivalent to primary formazin. The PCAL suspensions are used to calibrate the instrument response with respect to analyte concentration.
- 3.7 QUALITY CONTROL SAMPLE (QCS) -- A solution of the method analyte of known concentrations that is used to fortify an aliquot of LRB matrix. The QCS is obtained from a source external to the laboratory, and is used to check laboratory performance.
- 3.8 SECONDARY CALIBRATION STANDARDS (SCAL) -- Commercially prepared, stabilized sealed liquid or gel turbidity standards calibrated against properly prepared and diluted formazin, stabilized formazin or styrene divinylbenzene polymers.
- 3.9 STOCK STANDARD SUSPENSION (SSS) -- A concentrated suspension containing the analyte prepared in the laboratory using assayed reference materials or purchased from a reputable commercial source. Stock standard suspension or pre-diluted stabilized Formazin suspensions can be used to prepare calibration suspensions and other needed suspensions.

4.0 INTERFERENCES

- 4.1 The presence of floating debris and coarse sediments which settle out rapidly will give low readings. Finely divided air bubbles can cause high readings.
- 4.2 The presence of true color, that is the color of water which is due to dissolved substances that absorb light, will cause turbidities to be low, although this effect is generally not significant with drinking waters.
- 4.3 Light absorbing materials such as activated carbon in significant concentrations can cause low readings.

5.0 SAFETY

- 5.1 The toxicity or carcinogenicity of each reagent used in this method has not been fully established. Each chemical should be regarded as a potential health hazard and exposure should be as low as reasonably achievable.
- 5.2 Each laboratory is responsible for maintaining a current awareness file of OSHA regulations regarding the safe handling of the chemicals specified in this method. A reference file of Material Safety Data Sheets (MSDS) should be made available to all personnel involved in the chemical analysis. The preparation of a formal safety plan is also advisable.
- 5.3 Refer to all Material Safety Data Sheets (MSDSs) prior to preparing or using standards and before calibrating or performing instrument maintenance."

6.0 EQUIPMENT AND SUPPLIES

- 6.1 The turbidimeter shall consist of a nephelometer, with light source for illuminating the sample, and one or more photo-electric detectors with a readout device to indicate the intensity of light scattered at right angles to the path of the incident light. The turbidimeter should be designed so that little stray light

reaches the detector in the absence of turbidity and should be free from significant drift after a short warm-up period.

- 6.2 Differences in physical design of turbidimeters will cause differences in measured values for turbidity, even though the same suspension is used for calibration. To minimize such differences, the following design criteria should be observed:

6.2.1 Tungsten Optics

- 6.2.1.1 Light source: Tungsten lamp operated at a color temperature between 2200-3000°K.
- 6.2.1.2 Distance traversed by incident light and scattered light within the sample tube: Total not to exceed 10 cm.
- 6.2.1.3 Detector: Centered at 90° to the incident light path and not to exceed $\pm 30^\circ$ from 90°. The detector, and filter system if used, shall have a spectral peak response between 400 and 600 nm.
- 6.2.1.4 Equipment: Examples of Hach Company's turbidimeters which meet or exceed these specifications are as follows; 2100A, 2100N, 2100AN, 1720C, and 1720D.

6.2.2 ISO 7027 Optics

- 6.2.2.1 Light source: (LED) operate at a wavelength of 860 ± 30 nm.
 - 6.2.2.1.1 Tungsten lamps fitted with monochromators and filters, diodes and lasers may be used as sources of monochromatic radiation.
- 6.2.2.2 There shall be no divergence from parallelism at the incident radiation and any convergence shall not exceed 1.5° .
- 6.2.2.3 Distance traversed by incident light and scattered light within the sample tube if needed: Total not to exceed 10 cm.
- 6.2.2.4 Detector: Centered at 90° to the incident light path and not to exceed $\pm 2.5^\circ$ from 90°. The detector, and filter system if used, shall have a spectral peak response between 860 ± 30 nm.
- 6.2.2.5 Equipment: Examples of Hach Company's turbidimeters which meet or exceed these specifications are as follows; 1720D/L, 2100N IS and 2100AN IS

- 6.3 The sensitivity of the instrument should permit detection of a turbidity difference of 0.01NTU or less in waters having turbidities less than 1 unit. The instrument should measure from 0 to 40 units turbidity. Several ranges may be necessary to obtain both adequate coverage and sufficient sensitivity for low turbidities.
- 6.4 The sample tubes to be used with the available instrument must be of clear, colorless glass or plastic. They should be kept scrupulously clean, both inside and out, and discarded when they become scratched or etched. A light coating of silicon oil may be used to mask minor imperfections in glass tubes. They must not be handled at all where the light strikes them, but should be provided with sufficient extra length,

or with a protective case, so that they may be handled. Tubes should be checked, indexed and read at the orientation that produces the lowest background blank value.

6.5 Balance -- Analytical, capable of accurately weighing to the nearest 0.0001 g.

6.6 Glassware -- Class A volumetric flasks and pipets as required.

7.0 REAGENTS AND STANDARDS

7.1 Reagent water, turbidity-free: Pass deionized distilled water through a 0.45 μ pore size membrane filter, if such filtered water shows a lower turbidity than unfiltered distilled water.

7.2 Stock standard suspension (Formazin):

7.2.1 Dissolve 1.00 g hydrazine sulfate, $(\text{NH}_2)_2 \bullet \text{H}_2\text{SO}_4$, (CASRN 10034-93-2) in reagent water and dilute to 100 mL in a volumetric flask.

7.2.2 Dissolve 10.00 g hexamethylenetetramine (CASRN 100-97-0) in reagent water and dilute to 100 mL in a volumetric flask. In a 100 mL volumetric flask, mix 5.0 mL of each solution (7.2.1 + 7.2.2). Allow to stand 24 hours at $25 \pm 3^\circ\text{C}$, then dilute to the mark with reagent water.

7.3 Primary Calibration Standards

7.3.1 Prepared Standards

7.3.1.1 Mix and dilute 10.00 mL of stock standard suspension (7.2) to 100 mL with reagent water. The turbidity of this suspension is defined as 40 NTU. For other values, mix and dilute portions of this suspension as required.

7.3.1.2 A new stock standard suspension (7.2) should be prepared each month. Primary calibration standards (7.3) should be prepared daily by dilution of the stock standard suspension.

7.3.2 Commercially Available Standards

7.3.2.1 Formazin in commercially prepared primary concentrated stock standard suspension (SSS) may be diluted and used as required. Dilute turbidity standards should be prepared daily.

7.3.2.2 Prediluted stabilized Formazin suspensions (StablCal™) primary standards are available for use in all instruments and require no preparation or dilution prior to use.

7.4 Secondary Standard

7.4.1 May be acceptable as a daily calibration check, but must be monitored on a routine basis for deterioration and replaced as required.

8.0 SAMPLE COLLECTION, PRESERVATION AND STORAGE

8.1 Samples should be collected in plastic or glass bottles. All bottles must be thoroughly cleaned and rinsed with turbidity free water. Volume collected should be sufficient to insure a representative sample, allow for replicate analysis (if required), and minimize waste disposal.

8.2 No chemical preservation is required. Cool sample to 4°C.

8.3 Samples should be analyzed as soon as possible after collection. If storage is required, samples maintained at 4°C may be held for up to 48 h.

9.0 QUALITY CONTROL

9.1 Each laboratory using this method is required to operate a formal quality control (QC) program. The minimum requirements of this program consist of an initial demonstration of laboratory capability and analysis of laboratory reagent blanks and other solutions as a continuing check on performance. The laboratory is required to maintain performance records that define the quality of data generated.

9.2 INITIAL DEMONSTRATION OF PERFORMANCE.

9.2.1 The initial demonstration of performance is used to characterize instrument performance (determination of LCRs and analysis of QCS).

9.2.2 Linear Calibration Range (LCR) -- The LCR must be determined initially and verified every 6 months or whenever a significant change in instrument response is observed or expected. The initial demonstration of linearity must use sufficient standards to insure that the resulting curve is linear. The verification of linearity must use a minimum of a blank and three standards. If any verification data exceeds the initial values by $\pm 10\%$, linearity must be reestablished. If any portion of the range is shown to be nonlinear, sufficient standards must be used to clearly define the nonlinear portion.

9.2.3 Quality Control Sample (QCS) -- When beginning the use of this method, on a quarterly basis or as required to meet data-quality needs, verify the calibration standards and acceptable instrument performance with the preparation and analysis of a QCS. If the determined concentrations are not within $\pm 10\%$ of the stated values, performance of the determinative step of the method is unacceptable. The source of the problem must be identified and corrected before continuing with on-going analyses.

9.3 ASSESSING LABORATORY PERFORMANCE

9.3.1 Laboratory Reagent Blank (LRB) -- The laboratory must analyze at least one LRB with each batch of samples. Data produced are used to assess contamination from the laboratory environment.

9.3.2 Instrument Performance Check Solution (IPC) -- For all determinations, the laboratory must analyze the IPC (a midrange check standard - Section 3.3) and a calibration blank immediately following daily calibration, after every tenth sample (or more frequently, if required) and at the end of the sample run. Analysis of the IPC solution and calibration blank immediately following calibration must verify that the instrument is within $\pm 10\%$ of calibration. Subsequent analyses of the IPC solution must verify the calibration is still within $\pm 10\%$. If the calibration cannot be verified within the specified limits, reanalyze the IPC solution. If the second analysis of the IPC solution confirms calibration to be outside the limits, sample analysis must be discontinued, the cause determined and/or in the case of drift the instrument recalibrated. All samples following the last acceptable IPC solution must be reanalyzed. The analysis data of the calibration blank and IPC solution must be kept on file with the sample analyses data. NOTE: Secondary calibration standards (SS) may also be used as the IPC.

9.3.3 Where additional reference materials such as Performance Evaluation samples are available, they should be analyzed to provide additional performance data. The analysis of reference samples is a valuable tool for demonstrating the ability to perform the method acceptably.

10.0 CALIBRATION AND STANDARDIZATION

- 10.1 Turbidimeter calibration: The manufacturer's operating instructions should be followed. Measure standards on the turbidimeter covering the range of interest. If the instrument is already calibrated in standard turbidity units, this procedure will check the accuracy of the calibration scales. At least one standard should be run in each instrument range to be used. Some instruments permit adjustments of sensitivity so that scale values will correspond to turbidities. Solid standards, such as those made of lucite blocks, should never be used due to potential calibration changes caused by surface scratches. If a pre-calibrated scale is not supplied, calibration curves should be prepared for each range of the instrument.

11.0 PROCEDURE

- 11.1 Turbidities less than 40 units: If possible, allow samples to come to room temperature before analysis. Mix the sample to thoroughly disperse the solids. Wait until air bubbles disappear then pour the sample into the turbidimeter tube. Read the turbidity directly from the instrument scale or from the appropriate calibration curve.
- 11.2 Turbidities exceeding 40 units: Dilute the sample with one or more volumes of turbidity-free water until the turbidity falls below 40 units. The turbidity of the original sample is then computed from the turbidity of the diluted sample and the dilution factor. For example, if 5 volumes of turbidity-free water were added to 1 volume of sample, and the diluted sample showed a turbidity of 30 units, then the turbidity of the original sample was 180 units.
- 11.2.1 Some turbidimeters are equipped with several separate scales. The higher scales are to be used only as indicators of required dilution volumes to reduce readings to less than 40 NTU.

NOTE 1: Comparative work performed in the Environmental Monitoring Systems Laboratory - Cincinnati (EMSL-Cincinnati) indicates a progressive error on sample turbidities in excess of 40 units.

12.0 DATA ANALYSIS AND CALCULATIONS

- 12.1 Multiply sample readings by appropriate dilution to obtain final reading.
- 12.2 Report results as follows:

<u>NTU</u>	<u>Record to Nearest:</u>
0.0 - 1.0	0.05
1 -10	0.1
10 - 0	1
40 -100	5
100 - 400	10
400 - 1000	50
> 1000	100

13.0 METHOD PERFORMANCE

- 13.1 In a single laboratory (EMSL-Cincinnati), using surface water samples at levels of 26, 41, 75 and 180 NTU, the standard deviations were ± 0.60 , ± 0.94 , ± 1.2 and ± 4.7 units, respectively.

- 13.2 The interlaboratory precision and accuracy data in Table I were developed using a reagent water matrix. Values are in NTU. The intralaboratory precision and accuracy data in Table II were developed using various Hach instruments.

14.0 POLLUTION PREVENTION

- 14.1 Pollution prevention encompasses any technique that reduces or eliminates the quantity or toxicity of waste at the point of generation. Numerous opportunities for pollution prevention exist in laboratory operation. The EPA has established a preferred hierarchy of environmental management techniques that places pollution prevention as the management option of first choice.

Whenever feasible, laboratory personnel should use pollution prevention techniques to address their waste generation. When wastes cannot be feasibly reduced at the source, the Agency recommends recycling as the next best option.

- 14.2 The quantity of chemicals purchased should be based on expected usage during its shelf life and disposal cost of unused material. Actual reagent preparation volumes should reflect anticipated usage and reagent stability.
- 14.3 For information about pollution prevention that may be applicable to laboratories and research institutions, consult "Less is Better: Laboratory Chemical Management for Waste Reduction," available from the American Chemical Society's Department of Government Regulations and Science Policy, 1155 16th Street N.W., Washington D.C. 20036, (202)872-4477.

15.0 WASTE MANAGEMENT

- 15.1 The U.S. Environmental Protection Agency requires that laboratory waste management practices be conducted consistent with all applicable rules and regulations. Excess reagents, samples and method process wastes should be characterized and disposed of in an acceptable manner. The Agency urges laboratories to protect the air, water and land by minimizing and controlling all releases from hoods, and bench operations, complying with the letter and spirit of any waste discharge permit and regulations, and by complying with all solid and hazardous waste regulations, particularly the hazardous waste identification rules and land disposal restrictions.

For further information on waste management consult the "Waste Management Manual for Laboratory Personnel," available from the American Chemical Society at the address listed in Sect. 14.3.

16.0 REFERENCES

1. Annual Book of ASTM Standards, Volume 11.01 Water (1), Standard D1889-88A, p. 359, (1993).
2. Standard Methods for the Examination of Water and Wastewater, 18th Edition, pp. 2-9, Method 2130B, (1992).
3. Sadar, Michael J., StablCal™ Standardized Formazin Turbidity Standards, Lit No. 9581, Hach Company 1996

17.0 TABLES, DIAGRAMS, FLOWCHARTS AND VALIDATION DATA

TABLE I

INTERLABORATORY PRECISION AND ACCURACY DATA

Number Of Values Reported	True Value (T)	Mean (X)	Residual For X	Standard Deviation (S)	Residual For (S)
373	0.450	0.4864	0.0027	0.1071	-0.0078
374	0.600	0.6026	-0.0244	0.1048	-0.0211
289	0.65	0.6931	0.0183	0.1301	0.0005
482	0.910	0.9244	0.0013	0.2512	0.1024
484	0.910	0.9919	0.0688	0.1486	-0.0002
489	1.00	0.9405	-0.0686	0.1318	-0.0236
640	1.36	1.3456	-0.0074	0.1894	0.0075
487	3.40	3.2616	-0.0401	0.3219	-0.0103
288	4.8	4.5684	-0.0706	0.3776	-0.0577
714	5.60	5.6984	0.2952	0.4411	-0.0531
641	5.95	5.6026	-0.1350	0.4122	-0.1078

REGRESSIONS: $X = 0.955T + 0.54$, $S = 0.074T + 0.082$

TABLE II

INTRALABORATORY PRECISION AND ACCURACY DATA

Standard* (NTU)	2100AN	2100AN IS	2100A	1720D/L	1720D	1720C
0.139	0.137	0.121	0.16	0.128	0.133	0.148
0.30	0.309	0.289	0.33	0.301	0.314	0.309
0.50	0.512	0.498	0.52	0.495	0.493	0.527
1.00	1.050	1.035	1.01	1.047	1.036	1.079

Filtered Effluent Samples	2100AN	2100AN IS	2100A	1720D/L	1720D	1720C
#1	0.158	0.161	0.15	0.138	0.141	0.165
#2	0.046	0.033	0.06	0.045	0.046	0.071
#3	0.069	0.053	0.08	0.057	0.063	0.069
WS40 #2	0.335	0.312	0.33	n/a	0.320	n/a
WS40 #3	0.338	0.313	0.36	n/a	n/a	n/a
WS41 #2	0.206	0.188	n/a	n/a	0.211	n/a
WS41 #3	0.205	0.191	n/a	n/a	n/a	n/a
WS41 #4	0.175	0.163	n/a	n/a	n/a	n/a

* = All standards listed are ± 0.05 NTU except for 0.139 which is $\pm 5\%$ relative standard deviation.

Note:

All WS samples are from USEPA Water Supply Performance Evaluation Studies 40 and 41. Instruments with N/A indicate that not enough sample was available for collecting data.